

## INFORMED CONSENT: MICROCURRENT

### INTENDED USE

The bt-nano Facial Toning System is intended to stimulate the face. The device is indicated for cosmetic use.

Caution: Federal Law restricts this device to sale by or on the order of a physician or licensed practitioner.

### CONTRAINDICATIONS

Consult a physician regarding any questions on contraindications or other medical concerns. Do not use if any contraindication is present:

1. Pacemaker or other implanted electronic devices
2. Epilepsy
3. Pregnancy
4. Cancer
5. Thrombosis
6. Phlebitis
7. Under a physician's care for a disorder, illness, disease, or injury.

•The bt-nano Facial Toning System should be used with caution for undiagnosed pain syndromes until etiology has been established.

•Piercings of any kind must be removed prior to treatment, as they may provide a conductive path to areas that are sensitive and/or not intended to be treated

### PRECAUTIONS

The bt-nano is for use on healthy faces by healthy people. Consult your physician before use if this is not the case.

Headaches and other painful sensations have been reported during use or following electrical stimulation to the head and face. Lowering the intensity may reduce or eliminate these sensations. There also may be a perception of flashing lights during use, resulting from the stimulation of your optic nerve. If you notice this condition constantly and without the use of the bt-nano, please consult your doctor.

\_\_\_\_\_ I have read and acknowledged all statements on this page & give my consent to this (INT) treatment.

\_\_\_\_\_  
(PRINT NAME)

\_\_\_\_\_  
(SIGNATURE)

\_\_\_\_\_  
(DATE)